

Product datasheet

Human Anti-Phospholipid Antibody IgG ELISA Kit

Catalog # EK00838E

Overview

Product name	Human Anti-Phospholipid Antibody IgG (APL/APL/APA IgG) ELISA Kit
Size	96 wells
Species Reactivity	Human
Assay Type	Sandwich
Sample Type	Serum, Plasma, Other human body fluids & secretions
Assay Range	20U/mL – 1400U/mL
Kit Components	Assay Plate, Standard, Standard diluent, Sample diluent, Enzyme conjugate, Detection reagent A & B, Wash buffer, Stop solution, Protocol, Quality control certificate
Storage	Store at 2-8 °C
Product Note	<p>This ELISA Kit uses the double-antigen sandwich method to determine the level of human anti-phospholipid antibody IgG (APL/APL IgG) in the specimen. The ELISA analytical biochemical technique of the kit is based on phospholipid antigen-APL/APA IgG antibody interactions (immunosorbency) and an HRP colorimetric detection system to detect APL/APA IgG targets in samples. APL/APA IgG are added to the antigen-coated microwells, and then combine with the HRP-labeled antigen to form an antigen-antibody -Enzyme-labeled antigen complex, after thorough washing, add substrate TMB to develop color. TMB is converted into blue under the catalysis of HRP enzyme, and into the final yellow under the action of acid. The color intensity is positively correlated with APL/APA IgG in the sample. Absorbance (OD value) is measured at a wavelength of 450nm with a microplate reader, and the concentration of human APL/APA IgG in the sample can be calculated from the standard curve.</p>



www.leadingbiology.com



TEL:1-661-524-0262



2600 Hilltop DR BLDG B SUITE C138 Richmond CA 94806,USA

Typical Data

Results of a typical standard curve are provided **with the product** for **demonstration only** and should not be used to obtain test results.

A standard curve must be run for each set of samples assayed.

IMPORTANCE

After receiving the kit, please check the intactness of outer packaging, check whether the reagent bottles are broken, leaking or the amount of liquid is significantly reduced, etc. If the diluent is polluted such as flocculent bacterial colonies; the substrate is not transparent and the microtiter plate is leaking, etc; please contact the distributor. In rare cases, crystals may appear in the substrate. Normally the crystals will dissolve at room temperature, this will not affect the testing results.

